TECHNICIAN PRACTICAL PERFORMANCE TEST FEMTOSECOND LASER PROCESSING

Aj	pplicant's N	Name:			
Е	ye Bank:				
	xaminer's N lease print				
To	otal Points:	Aseptic Technic	ique Surgical Technique 2 - 11 Pass		
As Su	septic Poin argical Poi	ts Range: 12 nts Range: 18	2 - 11 Pass 10 or Less Fail 8 - 14 Pass 13 or Less Fail		
AS	<u>SEPTI</u>	C TECHN	NIQUE_		
1.	Appropriate, monitored environment for performing processing with femtosecond laser (e.g. records of routine culturing, acceptable CFU levels).				
	□1 □0	Yes No	If no, describe deficiency of environment		
2.			s the technician wearing moisture impermeable clothing with sterile sleeves/sterile p, mask and protective eye wear?)		
	□1 □0	Yes No	If no, what was technician wearing?		
3.	Wrapp	oing of instrumen	nts (includes double wrapped, single-wrapped, and visi-peel bag)		
	□1 □0	Acceptable Unacceptable	If unacceptable, describe:		
1.	Kit un	wrapped so that t	the sterility of instruments, equipment, and sterile field are not compromised		
	□1 □0	Yes No	If aseptic technique violated, note how:		
5.	Placen	nent of non-steril	le items around sterile field		
	□1 □0	Acceptable Unacceptable	If unacceptable, describe:		
5.	_		is (via 3-5 minute scrub or EtOH-based surgical hand rub product) per AORN s and Eye Bank's SOP		
	□1 □0	Appropriate	technique If not appropriate, describe		

7.	Gloving technique				
	□1 □0	Acceptable Unacceptable	If unacceptable, describe:		
8.	Transfer of corneoscleral disc or whole eye from container to sterile field				
	□1 □0	Acceptable Unacceptable	If unacceptable, describe:		
9.	Placement of instrumentation and equipment during procedure				
	□1 □0	Acceptable Unacceptable	If unacceptable, describe:		
10.	Instruments used for conjunctival resection- Per Eye Bank's SOP				
	□1 □0 □0 □1		If unacceptable, describe: SOP requires removal however technician did not perform 's SOP does not require		
11.	Measurement of the stromal bed while maintaining aseptic technique (ie. using pachymetry during procedure) – Per Eye Bank's SOP.				
	□1 □0 □0 □1	Not performed:	If unacceptable, describe: SOP requires, however technician did not perform e does not require (e.g. full thickness graft vs. DSAEK)		
12.	Transf	Transfer of corneoscleral disc to vial or viewing chamber of storage solution			
	□1 □0	Acceptable Unacceptable	If unacceptable, describe:		
		AL TECH			
1.		second Laser warn e Bank's SOP	n up and appropriate check(s) run according to manufacturer's recommendations –		
	□1 □0 □0	Acceptable Unacceptable Not performed:	If unacceptable, describe:SOP requires, however technician did not perform		
2.	Objective(s) for graft configuration established and reviewed. Laser settings programmed accordingly. – Per Eye Bank's SOP				
	□1 □0 □0	Acceptable Unacceptable Not performed:	If unacceptable, describe: SOP requires, however technician did not perform		

3.	Conjun	Conjunctiva removal (360° peritomy) – Per Eye Bank's SOP				
	□1 □0 □0 □1	Acceptable: complete removal Unacceptable: compromised cornea, incomplete removal or contaminated the tissue Not performed: SOP requires, however technician did not perform N/A: Eye Bank's SOP does not require				
4.	Cornea placed on Artificial Chamber satisfactorily (ie. centers cornea, does not fold, induce stress lines, or traumatize endothelium) – Per Eye Bank's SOP					
	□2 □1 □0 □0	Acceptable Acceptable: slight problems Unacceptable: many problems If many problems, describe: Not performed: SOP requires, however technician did not perform				
5.	Adequately assesses pressure (ie. does not induce stress lines or traumatize endothelium) – Per Eye Bank's SOP					
	□2 □1 □0 □0	Acceptable Acceptable: slight problems Unacceptable: many problems If many problems, describe: Not performed: SOP requires, however technician did not perform				
6.	Appropriate applanation of femotsecond laser interface onto the cornea.					
	□2 □1 □0	Acceptable Acceptable: slight problems Unacceptable: many problems If many problems, describe:				
7.	Anterio	Anterior Chamber Maintained				
	□2 □0	Acceptable Unacceptable: anterior chamber collapsed or high pressure fluid rupture through laser cut				
8.	Thickness and diameter of stromal bed measured - per Eye Bank's SOP (*Note: The timing of this step may vary between banks, but should occur at the time specified by each Eye Bank's SOP)					
	□1 □0 □0 □1	Acceptable Unacceptable Not performed: SOP requires, however technician did not perform N/A: Procedure does not require (e.g. full thickness graft vs. DSAEK)				
9.	Cornea removed from Artificial Chamber satisfactorily (ie. does not fold, induce stress lines, or traumatize endothelium) – Per Eye Bank's SOP					
	□2 □1 □0	Acceptable Acceptable: slight problems Unacceptable: many problems If many problems, describe:				
10.	Manual dexterity during procedure					
	□2 □1 □0	Acceptable Acceptable: slight problems Unacceptable: many problems If many problems describe:				

11.	Question: Did the technician perform this procedure as described in the Eye Bank policies and procedures manual?				
	□1 □0	Yes No			
12.	Question: Did the technician explain deviations, if any, from the procedure as described in the Eye Bank policies and procedures manual?				
	□1 □0	Yes (or not applicable) No			